

Unpacking

Unwrap all packages carefully and compare contents with the packing list, making sure all items arrived. If any part is missing, contact your local Hoefer, Inc. sales office. Inspect all components for damage that may have occurred while the unit was in transit. If any part appears damaged, contact the carrier immediately. Be sure to keep all packing material for damage claims or to use should it become necessary to return the unit.

Specifications

This declaration of conformity is only valid for the instrument when it is:

- used in laboratory locations,
- used as delivered from Hoefer, Inc. except for alterations described in the user manual, and
- connected to other CE-labeled instruments or products recommended or approved by Hoefer, Inc.

Max. voltage	200 V
Max. wattage	20 W
Max. amperage	100 mA
Max. operating temp.	45 °C
Max. buffer volume	1.2 liters
Gel size	15 cm wide × 10, 15, or 20 cm long
Environmental operating conditions	Indoor use: 4–40 °C Humidity up to 80% Altitude up to 2000 m Installation category II Pollution degree 2
Dimensions (w × l × d) (includes electrode posts)	18.2 × 36 × 14 cm (7.2 × 14.2 × 5.5 in.)
Weight (base, lid, and leads only)	0.82 kg (1.8 lb)
Product certifications	EN61010–1, UL61010A-1, CSA C22.2 1010.1, CE Certified